

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070819\
 Data File : FG005378.D
 Signal(s) : FID1A.ch
 Acq On : 08 Jul 2019 23:06
 Operator : DD\AJ
 Sample : 20 TRPH STD
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 09 07:56:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070819.M
 Quant Title :
 QLast Update : Tue Jul 09 07:54:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.774	2114187	19.443 ug/ml
Target Compounds			
1) N-OCTANE	2.355	2387988	19.296 ug/ml
2) N-DECANE	4.471	2437498	19.247 ug/ml
3) N-DODECANE	6.500	2476709	19.259 ug/ml
4) N-TETRADECANE	8.302	2478960	19.289 ug/ml
5) N-HEXADECANE	9.909	2487798	19.246 ug/ml
6) N-OCTADECANE	11.356	2459825	19.279 ug/ml
7) N-EICOSANE	12.671	2469967	19.333 ug/ml
8) N-DOCOSANE	13.873	2430269	19.333 ug/ml
10) N-TETRACOSANE	14.979	2399482	19.411 ug/ml
11) N-HEXACOSANE	16.003	2360429	19.481 ug/ml
12) N-OCTACOSANE	16.954	2338449	19.687 ug/ml
13) N-TRIACONTANE	17.844	2299008	19.854 ug/ml
14) N-DOTRIACONTANE	18.678	2172377	19.845 ug/ml
15) N-TETRATRIACONTANE	19.464	2072012	19.778 ug/ml
16) N-HEXATRIACONTANE	20.207	1911247	19.884 ug/ml
17) N-OCTATRIACONTANE	20.968	1747470	19.832 ug/ml
18) N-TETRACONTANE	21.929	1752716	19.911 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070819\
Data File : FG005378.D
Signal(s) : FID1A.ch
Acq On : 08 Jul 2019 23:06
Operator : DD\AJ
Sample : 20 TRPH STD
Misc :
ALS Vial : 5 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 09 07:56:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070819.M
Quant Title :
QLast Update : Tue Jul 09 07:54:02 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

