

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070819\
 Data File : FG005380.D
 Signal(s) : FID1A.ch
 Acq On : 09 Jul 2019 00:05
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 09 07:57:18 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070819.M
 Quant Title :
 QLast Update : Tue Jul 09 07:54:02 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.772	538789	4.955 ug/ml
Target Compounds			
1) N-OCTANE	2.356	601127	4.857 ug/ml
2) N-DECANE	4.470	608299	4.803 ug/ml
3) N-DODECANE	6.499	621657	4.834 ug/ml
4) N-TETRADECANE	8.301	625044	4.864 ug/ml
5) N-HEXADECANE	9.907	629747	4.872 ug/ml
6) N-OCTADECANE	11.355	618101	4.844 ug/ml
7) N-EICOSANE	12.669	629873	4.930 ug/ml
8) N-DOCOSANE	13.871	617001	4.908 ug/ml
10) N-TETRACOSANE	14.977	612115	4.952 ug/ml
11) N-HEXACOSANE	16.001	607203	5.011 ug/ml
12) N-OCTACOSANE	16.951	630311	5.306 ug/ml
13) N-TRIACONTANE	17.841	657144	5.675 ug/ml
14) N-DOTRIACONTANE	18.675	654377	5.978 ug/ml
15) N-TETRATRIACONTANE	19.460	561409	5.359 ug/ml
16) N-HEXATRIACONTANE	20.204	580397	6.038 ug/ml
17) N-OCTATRIACONTANE	20.966	456616	5.182 ug/ml
18) N-TETRACONTANE	21.923	554431	6.298 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070819\
Data File : FG005380.D
Signal(s) : FID1A.ch
Acq On : 09 Jul 2019 00:05
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 09 07:57:18 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070819.M
Quant Title :
QLast Update : Tue Jul 09 07:54:02 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

