

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070822\
 Data File : FG010128.D
 Signal(s) : FID1A.ch
 Acq On : 08 Jul 2022 13:48
 Operator : AJ\MA
 Sample : N3214-12
 Misc : 1 PPM WATER MDL
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 09 03:14:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
 Quant Title :
 QLast Update : Thu Jun 30 04:15:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.808	1866521	15.697 ug/ml
Target Compounds			
1) N-OCTANE	1.830	98407	0.792 ug/ml
2) N-DECANE	4.336	95038	0.748 ug/ml
3) N-DODECANE	6.514	89221	0.690 ug/ml
4) N-TETRADECANE	8.343	88050	0.668 ug/ml
5) N-HEXADECANE	9.951	91946	0.689 ug/ml
6) N-OCTADECANE	11.394	93741	0.693 ug/ml
7) N-EICOSANE	12.705	97657	0.729 ug/ml
8) N-DOCOSANE	13.905	100682	0.752 ug/ml
10) N-TETRACOSANE	15.011	99711	0.752 ug/ml
11) N-HEXACOSANE	16.033	100222	0.764 ug/ml
12) N-OCTACOSANE	16.984	103548	0.806 ug/ml
13) N-TRIACONTANE	17.872	110936	0.882 ug/ml
14) N-DOTRIACONTANE	18.706	103409	0.874 ug/ml
15) N-TETRATRIACONTANE	19.492	81963	0.771 ug/ml
16) N-HEXATRIACONTANE	20.234	59696	0.670 ug/mlm
17) N-OCTATRIACONTANE	20.946	57173	0.734 ug/mlm
18) N-TETRACONTANE	21.794	57076	0.728 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG070822\
Data File : FG010128.D
Signal(s) : FID1A.ch
Acq On : 08 Jul 2022 13:48
Operator : AJ\MA
Sample : N3214-12
Misc : 1 PPM WATER MDL
ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 09 03:14:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
Quant Title :
QLast Update : Thu Jun 30 04:15:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

