

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071020\
 Data File : FG007051.D
 Signal(s) : FID1A.ch
 Acq On : 10 Jul 2020 13:15
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 11 00:34:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG061520.M
 Quant Title :
 QLast Update : Tue Jun 16 03:44:57 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.794	6816756	52.819 ug/ml
Target Compounds			
2) N-DECANE	4.500	8295133	57.166 ug/ml
3) N-DODECANE	6.528	8240591	56.428 ug/ml
4) N-TETRADECANE	8.328	8116641	55.491 ug/ml
5) N-HEXADECANE	9.932	8047450	54.588 ug/ml
6) N-OCTADECANE	11.378	8073083	53.939 ug/ml
7) N-EICOSANE	12.692	7917581	53.159 ug/ml
8) N-DOCOSANE	13.893	7789729	52.518 ug/ml
10) N-TETRACOSANE	14.999	7951304	44.621 ug/ml
11) N-HEXACOSANE	16.023	7634241	49.117 ug/ml
12) N-OCTACOSANE	16.975	7722289	47.698 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071020\
Data File : FG007051.D
Signal(s) : FID1A.ch
Acq On : 10 Jul 2020 13:15
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 11 00:34:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG061520.M
Quant Title :
QLast Update : Tue Jun 16 03:44:57 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

