

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071520\
 Data File : FG007064.D
 Signal(s) : FID1A.ch
 Acq On : 15 Jul 2020 16:48
 Operator : DD\AJ
 Sample : 100 TRPH STD
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 16 02:21:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071520.M
 Quant Title :
 QLast Update : Thu Jul 16 02:20:51 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.798	16569198	98.019 ug/ml
Target Compounds			
1) N-OCTANE	2.384	17289980	93.292 ug/ml
2) N-DECANE	4.501	17755611	93.783 ug/ml
3) N-DODECANE	6.530	18000564	94.426 ug/ml
4) N-TETRADECANE	8.331	18130911	95.109 ug/ml
5) N-HEXADECANE	9.936	18359928	95.888 ug/ml
6) N-OCTADECANE	11.381	18788372	96.638 ug/ml
7) N-EICOSANE	12.694	18788672	97.240 ug/ml
8) N-DOCOSANE	13.897	18810472	97.725 ug/ml
10) N-TETRACOSANE	15.004	18917905	98.051 ug/ml
11) N-HEXACOSANE	16.026	18609244	97.955 ug/ml
12) N-OCTACOSANE	16.978	18400736	97.354 ug/ml
13) N-TRIACONTANE	17.867	18063923	95.677 ug/ml
14) N-DOTRIACONTANE	18.699	16628645	94.269 ug/ml
15) N-TETRATRIACONTANE	19.482	13536745	93.463 ug/ml
16) N-HEXATRIACONTANE	20.223	10081469	92.368 ug/ml
17) N-OCTATRIACONTANE	20.986	7929243	91.738 ug/ml
18) N-TETRACONTANE	21.950	7167314	90.790 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071520\
Data File : FG007064.D
Signal(s) : FID1A.ch
Acq On : 15 Jul 2020 16:48
Operator : DD\AJ
Sample : 100 TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 16 02:21:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071520.M
Quant Title :
QLast Update : Thu Jul 16 02:20:51 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

