

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071520\
 Data File : FG007066.D
 Signal(s) : FID1A.ch
 Acq On : 15 Jul 2020 17:50
 Operator : DD\AJ
 Sample : 20 TRPH STD
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 16 02:22:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071520.M
 Quant Title :
 QLast Update : Thu Jul 16 02:20:51 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.790	3199959	18.930 ug/ml
Target Compounds			
1) N-OCTANE	2.384	3341020	18.027 ug/ml
2) N-DECANE	4.497	3427904	18.106 ug/ml
3) N-DODECANE	6.523	3463214	18.167 ug/ml
4) N-TETRADECANE	8.323	3487687	18.295 ug/ml
5) N-HEXADECANE	9.928	3537212	18.474 ug/ml
6) N-OCTADECANE	11.374	3620082	18.620 ug/ml
7) N-EICOSANE	12.686	3623240	18.752 ug/ml
8) N-DOCOSANE	13.887	3630746	18.863 ug/ml
10) N-TETRACOSANE	14.993	3656630	18.952 ug/ml
11) N-HEXACOSANE	16.016	3615693	19.032 ug/ml
12) N-OCTACOSANE	16.967	3616038	19.132 ug/ml
13) N-TRIACONTANE	17.856	3659783	19.384 ug/ml
14) N-DOTRIACONTANE	18.690	3403483	19.295 ug/ml
15) N-TETRATRIACONTANE	19.476	2737273	18.899 ug/ml
16) N-HEXATRIACONTANE	20.217	2017564	18.485 ug/ml
17) N-OCTATRIACONTANE	20.980	1583559	18.321 ug/ml
18) N-TETRACONTANE	21.940	1395180	17.673 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071520\
Data File : FG007066.D
Signal(s) : FID1A.ch
Acq On : 15 Jul 2020 17:50
Operator : DD\AJ
Sample : 20 TRPH STD
Misc :
ALS Vial : 55 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 16 02:22:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071520.M
Quant Title :
QLast Update : Thu Jul 16 02:20:51 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

