

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071618\
 Data File : FG001680.D
 Signal(s) : FID1A.ch
 Acq On : 16 Jul 2018 14:39
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 17 08:09:38 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:11:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.512	5715816	49.129 ug/ml
Target Compounds			
1) N-OCTANE	2.177	6493896	50.788 ug/ml
2) N-DECANE	4.260	6549760	50.768 ug/ml
3) N-DODECANE	6.276	6587765	50.082 ug/ml
4) N-TETRADECANE	8.067	6611558	49.987 ug/ml
5) N-HEXADECANE	9.666	6650181	50.283 ug/ml
6) N-OCTADECANE	11.106	6683801	49.972 ug/ml
7) N-EICOSANE	12.415	6613799	49.547 ug/ml
8) N-DOCOSANE	13.613	6545007	49.186 ug/ml
10) N-TETRACOSANE	14.716	6434196	48.669 ug/ml
11) N-HEXACOSANE	15.737	6330368	48.083 ug/ml
12) N-OCTACOSANE	16.688	6208077	47.487 ug/ml
13) N-TRIACONTANE	17.576	6081587	46.678 ug/ml
14) N-DOTRIACONTANE	18.408	5860654	46.041 ug/ml
15) N-TETRATRIACONTANE	19.195	5760521	45.768 ug/ml
16) N-HEXATRIACONTANE	19.938	5458711	45.442 ug/ml
17) N-OCTATRIACONTANE	20.654	4806105	44.756 ug/ml
18) N-TETRACONTANE	21.501	4508579	43.229 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071618\
Data File : FG001680.D
Signal(s) : FID1A.ch
Acq On : 16 Jul 2018 14:39
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 17 08:09:38 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
Quant Title :
QLast Update : Mon Jul 02 14:11:21 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

