

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071621\
 Data File : FG008525.D
 Signal(s) : FID1A.ch
 Acq On : 16 Jul 2021 15:20
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 17 00:06:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071521.M
 Quant Title :
 QLast Update : Thu Jul 15 13:17:03 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.763	7609239	53.401 ug/ml
Target Compounds			
1) N-OCTANE	2.353	7970241	53.366 ug/ml
2) N-DECANE	4.463	8148532	54.091 ug/ml
3) N-DODECANE	6.491	8323662	54.086 ug/ml
4) N-TETRADECANE	8.293	8416313	54.020 ug/ml
5) N-HEXADECANE	9.899	8412888	54.187 ug/ml
6) N-OCTADECANE	11.345	8689222	53.742 ug/ml
7) N-EICOSANE	12.659	8718233	53.495 ug/ml
8) N-DOCOSANE	13.861	8692131	53.340 ug/ml
10) N-TETRACOSANE	14.968	8754678	53.380 ug/ml
11) N-HEXACOSANE	15.991	8675242	53.501 ug/ml
12) N-OCTACOSANE	16.944	8600991	53.671 ug/ml
13) N-TRIACONTANE	17.833	8534567	54.013 ug/ml
14) N-DOTRIACONTANE	18.668	8168792	54.374 ug/ml
15) N-TETRATRIACONTANE	19.451	7615833	55.265 ug/ml
16) N-HEXATRIACONTANE	20.193	6623686	56.211 ug/ml
17) N-OCTATRIACONTANE	20.948	5630322	56.435 ug/ml
18) N-TETRACONTANE	21.899	5067855	55.925 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071621\
Data File : FG008525.D
Signal(s) : FID1A.ch
Acq On : 16 Jul 2021 15:20
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 17 00:06:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071521.M
Quant Title :
QLast Update : Thu Jul 15 13:17:03 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

