

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071721\
 Data File : FG008543.D
 Signal(s) : FID1A.ch
 Acq On : 17 Jul 2021 04:02
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 17 05:59:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071521.M
 Quant Title :
 QLast Update : Thu Jul 15 13:17:03 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.760	7330757	51.447 ug/ml
Target Compounds			
2) N-DECANE	4.462	7730893	51.318 ug/ml
3) N-DODECANE	6.490	7893953	51.294 ug/ml
4) N-TETRADECANE	8.291	8008255	51.401 ug/ml
5) N-HEXADECANE	9.897	8010068	51.592 ug/ml
6) N-OCTADECANE	11.344	8286204	51.249 ug/ml
7) N-EICOSANE	12.657	8334211	51.139 ug/ml
8) N-DOCOSANE	13.858	8346056	51.216 ug/ml
10) N-TETRACOSANE	14.965	8400711	51.222 ug/ml
11) N-HEXACOSANE	15.988	8345735	51.469 ug/ml
12) N-OCTACOSANE	16.941	8288276	51.720 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071721\
Data File : FG008543.D
Signal(s) : FID1A.ch
Acq On : 17 Jul 2021 04:02
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 17 05:59:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG071521.M
Quant Title :
QLast Update : Thu Jul 15 13:17:03 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

