

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071818\
 Data File : FG001722.D
 Signal(s) : FID1A.ch
 Acq On : 18 Jul 2018 23:35
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 19 00:59:49 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:11:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.515	5409870	46.499 ug/ml
Target Compounds			
1) N-OCTANE	2.176	6100257	47.709 ug/ml
2) N-DECANE	4.260	6310741	48.915 ug/ml
3) N-DODECANE	6.276	6377632	48.484 ug/ml
4) N-TETRADECANE	8.069	6366168	48.132 ug/ml
5) N-HEXADECANE	9.668	6394222	48.348 ug/ml
6) N-OCTADECANE	11.109	6404978	47.887 ug/ml
7) N-EICOSANE	12.417	6314384	47.304 ug/ml
8) N-DOCOSANE	13.616	6204256	46.625 ug/ml
10) N-TETRACOSANE	14.719	6063117	45.862 ug/ml
11) N-HEXACOSANE	15.740	5936687	45.092 ug/ml
12) N-OCTACOSANE	16.690	5853845	44.777 ug/ml
13) N-TRIACONTANE	17.580	5856528	44.951 ug/ml
14) N-DOTRIACONTANE	18.412	5780563	45.412 ug/ml
15) N-TETRATRIACONTANE	19.199	5739198	45.598 ug/ml
16) N-HEXATRIACONTANE	19.941	5540509	46.123 ug/ml
17) N-OCTATRIACONTANE	20.657	5015764	46.708 ug/ml
18) N-TETRACONTANE	21.506	5068231	48.595 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG071818\
Data File : FG001722.D
Signal(s) : FID1A.ch
Acq On : 18 Jul 2018 23:35
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 19 00:59:49 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
Quant Title :
QLast Update : Mon Jul 02 14:11:21 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

