



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Ardmore Chemical

Project Name: PVSC Monthly 2025

Project # N/A

Order ID # Q2585

Test Name: SVOCMS Group1

A. Number of Samples and Date of Receipt:

1 Water sample was received on 07/11/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOCMS Group1. This data package contains results for SVOCMS Group1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of SVOCMS Group1 was based on method 625.1 and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for EFF-WW [2-Fluorophenol - 42%, Phenol-d6 - 26% and Terphenyl-d14 - 58%], due to matrix interference as evidenced by the chromatogram therefore no corrective action was taken.

The Internal Standards Areas were met for all analysis except for EFF-WW, due to matrix interference as evidenced by the chromatogram therefore no corrective action was taken. The Retention Times were met for all analysis.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank Spike Duplicate met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% for certain compounds in the Initial Calibration (Method 8270-BF071725.M) for 2,4-Dinitrophenol, 4,6-Dinitro-2-methylphenol these Compounds are passing on Linear Regression.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

E. Additional Comments:



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____