

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 07/24/2023
Supervised By :Ankita Jodhani 07/24/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012507.D
Data File : FG012507.D
Signal(s) : FID1A.ch
Acq On : 20 Jul 2023 16:40
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 53 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070623.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	4.521	4.478	4.595	BB	525225	5820593	99.37%	9.407%
2	6.691	6.648	6.747	BB	580810	5787853	98.81%	9.354%
3	8.520	8.429	8.581	BB	537638	5681529	97.00%	9.182%
4	10.130	10.082	10.202	BB	515360	5712811	97.53%	9.233%
5	11.574	11.524	11.613	BB	510468	5777124	98.63%	9.337%
6	12.886	12.840	12.942	VV	478293	5812329	99.23%	9.394%
7	14.086	14.049	14.118	PV	455751	5562843	94.97%	8.990%
8	14.989	14.950	15.050	PV	354844	4947509	84.47%	7.996%
9	15.193	15.105	15.235	PV	435358	5857418	100.00%	9.467%
10	16.215	16.176	16.253	M	429500	5499119	93.88%	8.887%
11	17.166	17.102	17.224	VV	405000	5415709	92.46%	8.753%
Sum of corrected areas:						61874837		

FG070623.M Fri Jul 21 04:32:52 2023