

rtres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012555.D
Data File : FG012555.D
Signal(s) : FID1A.ch
Acq On : 22 Jul 2023 01:55
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Reviewed By :Yogesh Patel 07/24/2023
Supervised By :Ankita Jodhani 07/24/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070623.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.006	1.980	2.103	BB	447764	6276350	98.33%	6.267%
2	4.515	4.471	4.588	BB	611837	6383070	100.00%	6.373%
3	6.686	6.643	6.745	BB	614851	6234012	97.66%	6.224%
4	8.514	8.468	8.597	BB	575012	6058542	94.92%	6.049%
5	10.124	10.073	10.208	BB	540076	6121847	95.91%	6.112%
6	11.568	11.472	11.625	BB	537448	6165390	96.59%	6.156%
7	12.879	12.838	12.949	BB	514886	6024256	94.38%	6.015%
8	14.079	14.043	14.133	VB	469223	5852373	91.69%	5.843%
9	14.979	14.929	15.006	BV	377408	5088500	79.72%	5.081%
10	15.184	15.105	15.264	BB	457390	6097786	95.53%	6.088%
11	16.205	16.128	16.268	BB	417690	5666742	88.78%	5.658%
12	17.157	17.003	17.213	BV	416065	5594380	87.64%	5.586%
13	18.046	17.974	18.099	BV	391644	5481617	85.88%	5.473%
14	18.879	18.754	18.932	VV	358919	5197415	81.43%	5.189%
15	19.663	19.619	19.701	PV	324603	4759113	74.56%	4.752%
16	20.405	20.336	20.445	BV	294978	4424047	69.31%	4.417%
17	21.140	21.085	21.185	VV	247038	4440959	69.57%	4.434%
18	22.046	21.988	22.104	PV	186268	4289848	67.21%	4.283%

Sum of corrected areas: 100156248

FG070623.M Sat Jul 22 02:37:51 2023