

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072318\
 Data File : FG001757.D
 Signal(s) : FID1A.ch
 Acq On : 23 Jul 2018 18:08
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jul 24 05:07:38 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:11:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.509	5989225	51.479 ug/ml
Target Compounds			
2) N-DECANE	4.255	7218444	55.951 ug/ml
3) N-DODECANE	6.272	7246318	55.088 ug/ml
4) N-TETRADECANE	8.065	7205207	54.475 ug/ml
5) N-HEXADECANE	9.663	7160598	54.142 ug/ml
6) N-OCTADECANE	11.103	7120588	53.238 ug/ml
7) N-EICOSANE	12.412	7000315	52.443 ug/ml
8) N-DOCOSANE	13.609	6875563	51.670 ug/ml
10) N-TETRACOSANE	14.713	6744264	51.015 ug/ml
11) N-HEXACOSANE	15.732	6621823	50.296 ug/ml
12) N-OCTACOSANE	16.682	6493001	49.666 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072318\
Data File : FG001757.D
Signal(s) : FID1A.ch
Acq On : 23 Jul 2018 18:08
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jul 24 05:07:38 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
Quant Title :
QLast Update : Mon Jul 02 14:11:21 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

