

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072324\
 Data File : FG013957.D
 Signal(s) : FID1A.ch
 Acq On : 23 Jul 2024 13:59
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Integration File: autoint1.e
 Quant Time: Jul 23 14:13:20 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
 Quant Title :
 QLast Update : Tue Jul 23 14:13:11 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.950	635707	5.255 ug/ml
Target Compounds			
1) N-OCTANE	1.971	649150	5.395 ug/ml
2) N-DECANE	4.474	644335	5.168 ug/ml
3) N-DODECANE	6.645	682588	5.173 ug/ml
4) N-TETRADECANE	8.477	688298	5.264 ug/ml
5) N-HEXADECANE	10.088	691703	5.190 ug/ml
6) N-OCTADECANE	11.533	715426	5.235 ug/ml
7) N-EICOSANE	12.847	700910	5.244 ug/ml
8) N-DOCOSANE	14.049	692316	5.259 ug/ml
10) N-TETRACOSANE	15.153	686911	5.262 ug/ml
11) N-HEXACOSANE	16.177	676186	5.264 ug/ml
12) N-OCTACOSANE	17.130	670587	5.287 ug/ml
13) N-TRIACONTANE	18.020	674313	5.361 ug/ml
14) N-DOTRIACONTANE	18.854	671686	5.505 ug/ml
15) N-TETRATRIACONTANE	19.640	541175	5.377 ug/ml
16) N-HEXATRIACONTANE	20.381	434218	5.197 ug/ml
17) N-OCTATRIACONTANE	21.114	420663	5.640 ug/ml
18) N-TETRACONTANE	22.015	366686	5.291 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072324\
Data File : FG013957.D
Signal(s) : FID1A.ch
Acq On : 23 Jul 2024 13:59
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
5 TRPH STD

Integration File: autoint1.e
Quant Time: Jul 23 14:13:20 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
Quant Title :
QLast Update : Tue Jul 23 14:13:11 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

