

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072518\
 Data File : FG001807.D
 Signal(s) : FID1A.ch
 Acq On : 26 Jul 2018 00:01
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 26 08:40:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:11:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.509	5652260	48.582 ug/ml
Target Compounds			
2) N-DECANE	4.255	7510561	58.215 ug/ml
3) N-DODECANE	6.272	7428038	56.470 ug/ml
4) N-TETRADECANE	8.064	7302054	55.208 ug/ml
5) N-HEXADECANE	9.663	7201158	54.449 ug/ml
6) N-OCTADECANE	11.103	7084750	52.970 ug/ml
7) N-EICOSANE	12.412	6852751	51.337 ug/ml
8) N-DOCOSANE	13.610	6619105	49.743 ug/ml
10) N-TETRACOSANE	14.713	6377571	48.241 ug/ml
11) N-HEXACOSANE	15.733	6192889	47.038 ug/ml
12) N-OCTACOSANE	16.683	6086809	46.559 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072518\
Data File : FG001807.D
Signal(s) : FID1A.ch
Acq On : 26 Jul 2018 00:01
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 26 08:40:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
Quant Title :
QLast Update : Mon Jul 02 14:11:21 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

