

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072522\
 Data File : FG010192.D
 Signal(s) : FID1A.ch
 Acq On : 25 Jul 2022 11:45
 Operator : YP\AJ
 Sample : N3789-02MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 IDW-VB1-071922MSD

Integration File: autoint1.e
 Quant Time: Jul 26 01:12:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
 Quant Title :
 QLast Update : Thu Jun 30 04:15:08 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.815	2263076	19.032 ug/ml
Target Compounds			
2) N-DECANE	4.338	2243011	17.644 ug/ml
3) N-DODECANE	6.517	2173653	16.817 ug/ml
4) N-TETRADECANE	8.348	2202303	16.702 ug/ml
5) N-HEXADECANE	9.957	2150950	16.127 ug/ml
6) N-OCTADECANE	11.401	2135592	15.794 ug/ml
7) N-EICOSANE	12.713	2158516	16.114 ug/ml
8) N-DOCOSANE	13.913	2167249	16.198 ug/ml
10) N-TETRACOSANE	15.019	2148109	16.207 ug/ml
11) N-HEXACOSANE	16.042	2078923	15.855 ug/ml
12) N-OCTACOSANE	16.994	2081246	16.204 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072522\
Data File : FG010192.D
Signal(s) : FID1A.ch
Acq On : 25 Jul 2022 11:45
Operator : YP\AJ
Sample : N3789-02MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
IDW-VB1-071922MSD

Integration File: autoint1.e
Quant Time: Jul 26 01:12:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG062922.M
Quant Title :
QLast Update : Thu Jun 30 04:15:08 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

