

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072718\
 Data File : FG001839.D
 Signal(s) : FID1A.ch
 Acq On : 27 Jul 2018 10:23
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 27 12:25:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:11:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.543	235027	2.020 ug/ml
Target Compounds				
2)	N-DECANE	4.253	7153470	55.447 ug/ml
3)	N-DODECANE	6.269	7081066	53.832 ug/ml
4)	N-TETRADECANE	8.060	6971941	52.712 ug/ml
5)	N-HEXADECANE	9.659	6877519	52.002 ug/ml
6)	N-OCTADECANE	11.099	6788989	50.759 ug/ml
7)	N-EICOSANE	12.407	6590960	49.376 ug/ml
8)	N-DOCOSANE	13.605	6400793	48.102 ug/ml
10)	N-TETRACOSANE	14.708	6228794	47.116 ug/ml
11)	N-HEXACOSANE	15.728	6041759	45.890 ug/ml
12)	N-OCTACOSANE	16.678	5960926	45.596 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072718\
Data File : FG001839.D
Signal(s) : FID1A.ch
Acq On : 27 Jul 2018 10:23
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 27 12:25:16 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
Quant Title :
QLast Update : Mon Jul 02 14:11:21 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

