

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG072718\
 Data File : FG001839.D
 Signal(s) : FID1A.ch
 Acq On : 27 Jul 2018 10:23
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 27 12:25:16 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:11:21 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
2	N-DECANE	129.015	143.069 E3	-10.9	112	-0.01
3	N-DODECANE	131.540	141.621 E3	-7.7	109	-0.02
4	N-TETRADECANE	132.265	139.439 E3	-5.4	106	-0.02
5	N-HEXADECANE	132.255	137.550 E3	-4.0	104	-0.02
6	N-OCTADECANE	133.751	135.780 E3	-1.5	102	-0.02
7	N-EICOSANE	133.485	131.819 E3	1.2	99	-0.02
8	N-DOCOSANE	133.066	128.016 E3	3.8	97	-0.02
9 S	TETRACOSANE-d50 (SURROGATE)	116.344	4.701 E3	96.0#	4#	0.02
10	N-TETRACOSANE	132.203	124.576 E3	5.8	95	-0.02
11	N-HEXACOSANE	131.656	120.835 E3	8.2	93	-0.02
12	N-OCTACOSANE	130.732	119.219 E3	8.8	92	-0.02

Evaluate Continuing Calibration Report - Not Found

1	N-OCTANE	127.863	0.000 E3	100.0#	0#	-2.18#
13	N-TRIACONTANE	130.287	0.000 E3	100.0#	0#	-17.59#
14	N-DOTRIACONTANE	127.293	0.000 E3	100.0#	0#	-18.42#
15	N-TETRATRIACONTANE	125.864	0.000 E3	100.0#	0#	-19.21#
16	N-HEXATRIACONTANE	120.126	0.000 E3	100.0#	0#	-19.95#
17	N-OCTATRIACONTANE	107.385	0.000 E3	100.0#	0#	-20.67#
18	N-TETRACONTANE	104.295	0.000 E3	100.0#	0#	-21.52#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0