

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG073018\
 Data File : FG001877.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jul 2018 13:01
 Operator : AJ\SJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 30 13:22:56 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG073018.M
 Quant Title :
 QLast Update : Mon Jul 30 13:20:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.470	1300357	10.488 ug/ml
Target Compounds			
1) N-OCTANE	2.155	1463303	10.107 ug/ml
2) N-DECANE	4.227	1489745	10.089 ug/ml
3) N-DODECANE	6.238	1524889	10.172 ug/ml
4) N-TETRADECANE	8.029	1550210	10.272 ug/ml
5) N-HEXADECANE	9.627	1555333	10.297 ug/ml
6) N-OCTADECANE	11.066	1578357	10.469 ug/ml
7) N-EICOSANE	12.375	1556167	10.531 ug/ml
8) N-DOCOSANE	13.572	1515428	10.526 ug/ml
10) N-TETRACOSANE	14.673	1481583	10.565 ug/ml
11) N-HEXACOSANE	15.695	1455106	10.605 ug/ml
12) N-OCTACOSANE	16.644	1442630	10.724 ug/ml
13) N-TRIACONTANE	17.533	1410485	10.804 ug/ml
14) N-DOTRIACONTANE	18.365	1275927	10.787 ug/ml
15) N-TETRATRIACONTANE	19.151	1090689	10.270 ug/ml
16) N-HEXATRIACONTANE	19.894	942602	9.893 ug/ml
17) N-OCTATRIACONTANE	20.605	816731	9.760 ug/ml
18) N-TETRACONTANE	21.441	800392	9.674 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG073018\
Data File : FG001877.D
Signal(s) : FID1A.ch
Acq On : 30 Jul 2018 13:01
Operator : AJ\SJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 30 13:22:56 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG073018.M
Quant Title :
QLast Update : Mon Jul 30 13:20:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

