

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG073018\
 Data File : FG001878.D
 Signal(s) : FID1A.ch
 Acq On : 30 Jul 2018 13:31
 Operator : AJ\SJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jul 30 13:56:38 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG073018.M
 Quant Title :
 QLast Update : Mon Jul 30 13:20:51 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.469	641350	5.110 ug/ml
Target Compounds			
1) N-OCTANE	2.156	708100	4.878 ug/ml
2) N-DECANE	4.227	724765	4.898 ug/ml
3) N-DODECANE	6.239	744575	4.945 ug/ml
4) N-TETRADECANE	8.029	763947	5.028 ug/ml
5) N-HEXADECANE	9.626	764571	5.025 ug/ml
6) N-OCTADECANE	11.065	784082	5.140 ug/ml
7) N-EICOSANE	12.373	777947	5.196 ug/ml
8) N-DOCOSANE	13.570	754668	5.174 ug/ml
10) N-TETRACOSANE	14.673	740041	5.204 ug/ml
11) N-HEXACOSANE	15.694	735957	5.284 ug/ml
12) N-OCTACOSANE	16.644	748614	5.466 ug/ml
13) N-TRIACONTANE	17.531	750662	5.637 ug/ml
14) N-DOTRIACONTANE	18.364	683924	5.671 ug/ml
15) N-TETRATRIACONTANE	19.149	579775	5.423 ug/ml
16) N-HEXATRIACONTANE	19.891	501342	5.276 ug/ml
17) N-OCTATRIACONTANE	20.606	442583	5.321 ug/ml
18) N-TETRACONTANE	21.438	432403	5.269 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG073018\
Data File : FG001878.D
Signal(s) : FID1A.ch
Acq On : 30 Jul 2018 13:31
Operator : AJ\SJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Jul 30 13:56:38 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG073018.M
Quant Title :
QLast Update : Mon Jul 30 13:20:51 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

