

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG073124\
 Data File : FG014004.D
 Signal(s) : FID1A.ch
 Acq On : 31 Jul 2024 19:45
 Operator : YP\AJ
 Sample : P3390-07
 Misc : 2.5 PPM WATER LOD
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2024

Integration File: autoint1.e
 Quant Time: Aug 01 03:03:53 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
 Quant Title :
 QLast Update : Tue Jul 23 14:43:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.949	1996957	16.509 ug/ml
Target Compounds			
2) N-DECANE	4.471	368151	2.953 ug/ml
3) N-DODECANE	6.643	387380	2.936 ug/ml
4) N-TETRADECANE	8.475	407214	3.114 ug/ml
5) N-HEXADECANE	10.086	397510	2.983 ug/ml
6) N-OCTADECANE	11.532	408998	2.993 ug/ml
7) N-EICOSANE	12.845	409922	3.067 ug/ml
8) N-DOCOSANE	14.046	406036	3.085 ug/ml
10) N-TETRACOSANE	15.152	398757	3.055 ug/ml
11) N-HEXACOSANE	16.177	393723	3.065 ug/ml
12) N-OCTACOSANE	17.128	394313	3.109 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG073124\
Data File : FG014004.D
Signal(s) : FID1A.ch
Acq On : 31 Jul 2024 19:45
Operator : YP\AJ
Sample : P3390-07
Misc : 2.5 PPM WATER LOD
ALS Vial : 67 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
LOD-MDL-WATER-01-QT3-2024

Integration File: autoint1.e
Quant Time: Aug 01 03:03:53 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
Quant Title :
QLast Update : Tue Jul 23 14:43:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

