

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080119\
 Data File : FG005553.D
 Signal(s) : FID1A.ch
 Acq On : 01 Aug 2019 09:42
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 01 14:03:09 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070819.M
 Quant Title :
 QLast Update : Tue Jul 09 08:00:47 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.767	5492126	51.030 ug/ml
Target Compounds			
2) N-DECANE	4.463	7354918	59.408 ug/ml
3) N-DODECANE	6.494	7395693	58.682 ug/ml
4) N-TETRADECANE	8.296	7250576	57.435 ug/ml
5) N-HEXADECANE	9.903	7110518	55.962 ug/ml
6) N-OCTADECANE	11.350	6841948	54.592 ug/ml
7) N-EICOSANE	12.665	6662015	52.800 ug/ml
8) N-DOCOSANE	13.866	6402951	51.641 ug/ml
10) N-TETRACOSANE	14.973	6238090	50.966 ug/ml
11) N-HEXACOSANE	15.996	6106436	50.820 ug/ml
12) N-OCTACOSANE	16.948	6072772	50.711 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080119\
Data File : FG005553.D
Signal(s) : FID1A.ch
Acq On : 01 Aug 2019 09:42
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 01 14:03:09 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG070819.M
Quant Title :
QLast Update : Tue Jul 09 08:00:47 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

