

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080124\
 Data File : FG014015.D
 Signal(s) : FID1A.ch
 Acq On : 01 Aug 2024 12:21
 Operator : YP\AJ
 Sample : P3390-06
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 02 05:34:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
 Quant Title :
 QLast Update : Tue Jul 23 14:43:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.948	1483822	12.267 ug/ml
Target Compounds			
1) N-OCTANE	1.965	151237	1.257 ug/ml
2) N-DECANE	4.469	157911	1.267 ug/ml
3) N-DODECANE	6.641	158734	1.203 ug/ml
4) N-TETRADECANE	8.474	178312	1.364 ug/ml
5) N-HEXADECANE	10.085	170333	1.278 ug/ml
6) N-OCTADECANE	11.530	174260	1.275 ug/ml
7) N-EICOSANE	12.843	175759	1.315 ug/ml
8) N-DOCOSANE	14.045	177479	1.348 ug/ml
10) N-TETRACOSANE	15.151	172338	1.320 ug/ml
11) N-HEXACOSANE	16.175	173810	1.353 ug/ml
12) N-OCTACOSANE	17.126	181443	1.431 ug/ml
13) N-TRIACONTANE	18.016	206041	1.638 ug/ml
14) N-DOTRIACONTANE	18.851	218584	1.791 ug/ml
15) N-TETRATRIACONTANE	19.637	186704	1.855 ug/ml
16) N-HEXATRIACONTANE	20.380	152708	1.828 ug/ml
17) N-OCTATRIACONTANE	21.111	131057	1.757 ug/ml
18) N-TETRACONTANE	22.012	115957	1.673 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080124\
Data File : FG014015.D
Signal(s) : FID1A.ch
Acq On : 01 Aug 2024 12:21
Operator : YP\AJ
Sample : P3390-06
Misc :
ALS Vial : 14 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 02 05:34:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
Quant Title :
QLast Update : Tue Jul 23 14:43:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

