

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080124\
 Data File : FG014020.D
 Signal(s) : FID1A.ch
 Acq On : 01 Aug 2024 15:19
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 02 05:35:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
 Quant Title :
 QLast Update : Tue Jul 23 14:43:06 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.951	5912289	48.876 ug/ml
Target Compounds			
2) N-DECANE	4.469	6433066	51.598 ug/ml
3) N-DODECANE	6.644	6603803	50.049 ug/ml
4) N-TETRADECANE	8.476	6406245	48.996 ug/ml
5) N-HEXADECANE	10.087	6565962	49.264 ug/ml
6) N-OCTADECANE	11.534	6674293	48.840 ug/ml
7) N-EICOSANE	12.847	6473791	48.438 ug/ml
8) N-DOCOSANE	14.050	6387074	48.520 ug/ml
10) N-TETRACOSANE	15.156	6342909	48.593 ug/ml
11) N-HEXACOSANE	16.178	6274275	48.849 ug/ml
12) N-OCTACOSANE	17.131	6275551	49.478 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080124\
Data File : FG014020.D
Signal(s) : FID1A.ch
Acq On : 01 Aug 2024 15:19
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 02 05:35:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072324.M
Quant Title :
QLast Update : Tue Jul 23 14:43:06 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

