

Instrument :

FID_G

LabSampleId :

10 TRPH STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012608.D

Data File : FG012608.D

Signal(s) : FID1A.ch

Acq On : 02 Aug 2023 17:55

Sample : 10 TRPH STD

Misc :

ALS Vial : 6 Sample Multiplier: 1

Reviewed By :Yogesh Patel 08/03/2023

Supervised By :Ankita Jodhani 08/03/2023

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080223.M

Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.080	2.044	2.124	BB	131302	1237252	93.82%	5.971%
2	4.596	4.567	4.626	BB	129467	1277550	96.88%	6.166%
3	6.769	6.738	6.810	BB	128709	1293648	98.10%	6.243%
4	8.601	8.562	8.627	BB	128235	1295304	98.23%	6.251%
5	10.213	10.183	10.258	BB	123535	1304557	98.93%	6.296%
6	11.660	11.619	11.703	BB	118621	1318701	100.00%	6.364%
7	12.974	12.929	13.018	BB	108657	1289537	97.79%	6.223%
8	14.175	14.130	14.220	BB	102919	1266363	96.03%	6.112%
9	15.075	15.032	15.128	BB	84553	1119262	84.88%	5.402%
10	15.280	15.236	15.314	BB	97092	1264692	95.90%	6.104%
11	16.303	16.257	16.350	BB	96687	1251803	94.93%	6.041%
12	17.253	17.203	17.285	BV	92985	1230029	93.28%	5.936%
13	18.140	18.091	18.191	BB	90519	1194235	90.56%	5.764%
14	18.973	18.925	19.046	BB	79365	1174195	89.04%	5.667%
15	19.757	19.707	19.797	BB	67285	972374	73.74%	4.693%
16	20.496	20.446	20.527	BV	56218	832532	63.13%	4.018%
17	21.251	21.185	21.312	BB	40293	738155	55.98%	3.562%
18	22.200	22.120	22.245	BV	28210	660456	50.08%	3.187%

Sum of corrected areas: 20720644

FG080223.M Thu Aug 03 01:42:42 2023