

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080322\
 Data File : FG010267.D
 Signal(s) : FID1A.ch
 Acq On : 03 Aug 2022 13:29
 Operator : YP\AJ
 Sample : N3980-04
 Misc : 2.5 PPM WATER LOD
 ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 03 20:50:28 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
 Quant Title :
 QLast Update : Thu Jul 28 15:10:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound		R.T.	Response	Conc Units

System Monitoring Compounds				
9) S	TETRACOSANE-d50 (SURR...	14.799	1860832	14.649 ug/ml
Target Compounds				
1)	N-OCTANE	1.826	267823	2.050 ug/ml
2)	N-DECANE	4.329	252542	1.880 ug/ml
3)	N-DODECANE	6.507	238880	1.738 ug/ml
4)	N-TETRADECANE	8.336	249552	1.781 ug/ml
5)	N-HEXADECANE	9.944	244381	1.690 ug/ml
6)	N-OCTADECANE	11.387	243705	1.652 ug/ml
7)	N-EICOSANE	12.699	253906	1.714 ug/ml
8)	N-DOCOSANE	13.897	259921	1.793 ug/ml
10)	N-TETRACOSANE	15.002	260247	1.804 ug/ml
11)	N-HEXACOSANE	16.024	254786	1.803 ug/ml
12)	N-OCTACOSANE	16.975	254083	1.840 ug/ml
13)	N-TRIACONTANE	17.863	250970	1.853 ug/ml
14)	N-DOTRIACONTANE	18.697	244943	1.865 ug/ml
15)	N-TETRATRIACONTANE	19.481	198956	1.590 ug/ml
16)	N-HEXATRIACONTANE	20.222	181971	1.578 ug/ml
17)	N-OCTATRIACONTANE	20.933	193791	1.705 ug/mlm
18)	N-TETRACONTANE	21.776	173402	1.478 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080322\
Data File : FG010267.D
Signal(s) : FID1A.ch
Acq On : 03 Aug 2022 13:29
Operator : YP\AJ
Sample : N3980-04
Misc : 2.5 PPM WATER LOD
ALS Vial : 56 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 03 20:50:28 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
Quant Title :
QLast Update : Thu Jul 28 15:10:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

