

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080322\
 Data File : FG010271.D
 Signal(s) : FID1A.ch
 Acq On : 03 Aug 2022 15:28
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 03 20:51:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
 Quant Title :
 QLast Update : Thu Jul 28 15:10:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.803	6228428	49.030 ug/ml
Target Compounds			
1) N-OCTANE	1.823	6447185	49.351 ug/ml
2) N-DECANE	4.329	6772161	50.418 ug/ml
3) N-DODECANE	6.510	6933515	50.443 ug/ml
4) N-TETRADECANE	8.339	7049098	50.318 ug/ml
5) N-HEXADECANE	9.947	7220720	49.934 ug/ml
6) N-OCTADECANE	11.391	7299212	49.482 ug/ml
7) N-EICOSANE	12.703	7268239	49.079 ug/ml
8) N-DOCOSANE	13.903	7077543	48.830 ug/ml
10) N-TETRACOSANE	15.008	7019814	48.665 ug/ml
11) N-HEXACOSANE	16.030	6863869	48.562 ug/ml
12) N-OCTACOSANE	16.981	6736706	48.789 ug/ml
13) N-TRIACONTANE	17.868	6652983	49.116 ug/ml
14) N-DOTRIACONTANE	18.701	6537701	49.766 ug/ml
15) N-TETRATRIACONTANE	19.486	6262541	50.040 ug/ml
16) N-HEXATRIACONTANE	20.226	5887782	51.059 ug/ml
17) N-OCTATRIACONTANE	20.941	5913895	52.035 ug/ml
18) N-TETRACONTANE	21.786	6319057	53.853 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080322\
Data File : FG010271.D
Signal(s) : FID1A.ch
Acq On : 03 Aug 2022 15:28
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 03 20:51:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
Quant Title :
QLast Update : Thu Jul 28 15:10:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

