

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080322\
 Data File : FG010273.D
 Signal(s) : FID1A.ch
 Acq On : 03 Aug 2022 16:28
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 03 21:07:02 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
 Quant Title :
 QLast Update : Thu Jul 28 15:10:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.803	6045881	47.593 ug/ml
Target Compounds			
2) N-DECANE	4.329	6721698	50.042 ug/ml
3) N-DODECANE	6.510	6880319	50.056 ug/ml
4) N-TETRADECANE	8.340	6979796	49.823 ug/ml
5) N-HEXADECANE	9.947	7132753	49.326 ug/ml
6) N-OCTADECANE	11.392	7189388	48.738 ug/ml
7) N-EICOSANE	12.703	7121187	48.086 ug/ml
8) N-DOCOSANE	13.902	6897558	47.588 ug/ml
10) N-TETRACOSANE	15.007	6816320	47.254 ug/ml
11) N-HEXACOSANE	16.029	6650068	47.049 ug/ml
12) N-OCTACOSANE	16.980	6519023	47.212 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080322\
Data File : FG010273.D
Signal(s) : FID1A.ch
Acq On : 03 Aug 2022 16:28
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 03 21:07:02 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
Quant Title :
QLast Update : Thu Jul 28 15:10:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

