

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080422\
 Data File : FG010295.D
 Signal(s) : FID1A.ch
 Acq On : 04 Aug 2022 15:22
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 05 05:55:46 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
 Quant Title :
 QLast Update : Thu Jul 28 15:10:44 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.802	6563719	51.670 ug/ml
Target Compounds			
2) N-DECANE	4.328	7243590	53.928 ug/ml
3) N-DODECANE	6.508	7419444	53.978 ug/ml
4) N-TETRADECANE	8.339	7532011	53.765 ug/ml
5) N-HEXADECANE	9.947	7709889	53.317 ug/ml
6) N-OCTADECANE	11.391	7792376	52.825 ug/ml
7) N-EICOSANE	12.701	7730022	52.197 ug/ml
8) N-DOCOSANE	13.902	7484147	51.635 ug/ml
10) N-TETRACOSANE	15.007	7386462	51.207 ug/ml
11) N-HEXACOSANE	16.028	7167649	50.711 ug/ml
12) N-OCTACOSANE	16.978	7012904	50.789 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080422\
Data File : FG010295.D
Signal(s) : FID1A.ch
Acq On : 04 Aug 2022 15:22
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 05 05:55:46 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG072822.M
Quant Title :
QLast Update : Thu Jul 28 15:10:44 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

