

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080624AR\
Data File : FG014072.D
Signal(s) : FID1A.ch
Acq On : 06 Aug 2024 18:45
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Aug 07 01:25:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aromatic EPH 080624.M
Quant Title : GC Extractables
QLast Update : Tue Aug 06 17:31:37 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm

Compound	R.T.	Response	Conc Units
----------	------	----------	------------

System Monitoring Compounds

Target Compounds

16) T	benzo[b]fluoranthene ...	16.665	98942	0.800 ug/ml
17) T	Bnezo[k]fluoranthene ...	16.665	98942	0.740 ug/ml
18) T	Benzo[a]pyrene (C31.34)	17.023	60693	0.481 ug/ml
19) T	Indeno[1,2,3-cd]pyren...	18.389	169294	1.654 ug/ml
20) T	Dibenz[a,h]anthracene...	18.430	379916	2.489 ug/ml
21) T	Benzo[g,h,i]perylene ...	18.639	255582	1.841 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080624AR\
Data File : FG014072.D
Signal(s) : FID1A.ch
Acq On : 06 Aug 2024 18:45
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :

Integration File: autoint1.e
Quant Time: Aug 07 01:25:29 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\Aromatic EPH 080624.M
Quant Title : GC Extractables
QLast Update : Tue Aug 06 17:31:37 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18µm






#4 2-Bromonaphthalene (SURRE)


R.T.: 0.000 min
Exp R.T. : 7.895 min
Response: 0
Conc: N.D.

Instrument :
FID_G
ClientSampleId :




#6 2-Fluorobiphenyl (SURRE)

R.T.: 0.000 min
Exp R.T. : 8.769 min
Response: 0
Conc: N.D.




#11 ortho-Terphenyl (SURRE)

R.T.: 0.000 min
Exp R.T. : 11.834 min
Response: 0
Conc: N.D.



#16 benzo[b]fluoranthene (C30.41)


R.T.: 16.665 min
Delta R.T.: 0.039 min
Response: 98942
Conc: 0.80 ug/ml



#17 Bnezo[k]fluoranthene (C30.14)


R.T.: 16.665 min
Delta R.T.: 0.007 min
Response: 98942
Conc: 0.74 ug/ml

Instrument :
FID_G
ClientSampleId :




#18 Benzo[a]pyrene (C31.34)

R.T.: 17.023 min
Delta R.T.: 0.013 min
Response: 60693
Conc: 0.48 ug/ml




#19 Indeno[1,2,3-cd]pyrene (C35.01)

R.T.: 18.389 min
Delta R.T.: 0.005 min
Response: 169294
Conc: 1.65 ug/ml



#20 Dibenz[a,h]anthracene (C30.36)

R.T.: 18.430 min
Delta R.T.: 0.004 min
Response: 379916
Conc: 2.49 ug/ml



#21 Benzo[g,h,i]perylene (C34.01)

R.T.: 18.639 min
Delta R.T.: 0.003 min
Response: 255582
Conc: 1.84 ug/ml

Instrument :

FID_G

ClientSampleId :