

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080819\
 Data File : FG005584.D
 Signal(s) : FID1A.ch
 Acq On : 08 Aug 2019 15:13
 Operator : DD\AJ
 Sample : 10 TRPH STD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 10 TRPH STD

Manual Integrations
 APPROVED

Ankita
 8/9/2019 4:02:45 PM

Integration File: autoint1.e
 Quant Time: Aug 08 15:31:43 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
 Quant Title :
 QLast Update : Thu Aug 08 15:05:10 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.756	1217147	9.241 ug/ml
Target Compounds			
1) N-OCTANE	2.346	1255853	8.544 ug/ml
2) N-DECANE	4.455	1279420	8.490 ug/ml
3) N-DODECANE	6.484	1324894	8.578 ug/ml
4) N-TETRADECANE	8.286	1349014	8.673 ug/ml
5) N-HEXADECANE	9.892	1385248	8.812 ug/ml
6) N-OCTADECANE	11.339	1393918	8.972 ug/ml
7) N-EICOSANE	12.652	1413849	9.090 ug/ml
8) N-DOCOSANE	13.854	1400739	9.184 ug/ml
10) N-TETRACOSANE	14.960	1385342	9.243 ug/ml
11) N-HEXACOSANE	15.983	1379689	9.375 ug/ml
12) N-OCTACOSANE	16.935	1500906	10.165 ug/ml
13) N-TRIACONTANE	17.824	1523853	10.584 ug/ml
14) N-DOTRIACONTANE	18.659	1302724	9.782 ug/ml
15) N-TETRATRIACONTANE	19.445	1180884	9.402 ug/mlm
16) N-HEXATRIACONTANE	20.187	1075408	9.062 ug/mlm
17) N-OCTATRIACONTANE	20.945	961982	8.940 ug/mlm
18) N-TETRACONTANE	21.895	933731	8.555 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080819\
Data File : FG005584.D
Signal(s) : FID1A.ch
Acq On : 08 Aug 2019 15:13
Operator : DD\AJ
Sample : 10 TRPH STD
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_G
ClientSampled :
10 TRPH STD

Manual Integrations
APPROVED

Ankita
8/9/2019 4:02:45 PM

Integration File: autoint1.e
Quant Time: Aug 08 15:31:43 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 15:05:10 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

