

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080819\
 Data File : FG005585.D
 Signal(s) : FID1A.ch
 Acq On : 08 Aug 2019 15:44
 Operator : DD\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Ankita
 8/9/2019 4:02:47 PM

Integration File: autoint1.e
 Quant Time: Aug 08 16:02:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
 Quant Title :
 QLast Update : Thu Aug 08 15:33:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.755	609788	4.719 ug/ml
Target Compounds			
1) N-OCTANE	2.346	649526	4.586 ug/ml
2) N-DECANE	4.455	657796	4.536 ug/ml
3) N-DODECANE	6.484	683225	4.587 ug/ml
4) N-TETRADECANE	8.286	698264	4.643 ug/ml
5) N-HEXADECANE	9.892	704401	4.618 ug/ml
6) N-OCTADECANE	11.338	702157	4.638 ug/ml
7) N-EICOSANE	12.652	709716	4.669 ug/ml
8) N-DOCOSANE	13.853	701266	4.694 ug/ml
10) N-TETRACOSANE	14.959	692573	4.710 ug/ml
11) N-HEXACOSANE	15.983	690994	4.770 ug/ml
12) N-OCTACOSANE	16.935	814006	5.490 ug/ml
13) N-TRIACONTANE	17.825	859634	5.885 ug/ml
14) N-DOTRIACONTANE	18.658	698610	5.275 ug/ml
15) N-TETRATRIACONTANE	19.444	613784	4.961 ug/mlm
16) N-HEXATRIACONTANE	20.189	564634	4.872 ug/mlm
17) N-OCTATRIACONTANE	20.947	515149	4.918 ug/mlm
18) N-TETRACONTANE	21.895	495775	4.712 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG080819\
Data File : FG005585.D
Signal(s) : FID1A.ch
Acq On : 08 Aug 2019 15:44
Operator : DD\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Ankita
8/9/2019 4:02:47 PM

Integration File: autoint1.e
Quant Time: Aug 08 16:02:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 15:33:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

