

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG081621\
 Data File : FG008657.D
 Signal(s) : FID1A.ch
 Acq On : 16 Aug 2021 15:45
 Operator : DD\AJ
 Sample : FG081621ICV
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 17 04:15:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG081621.M
 Quant Title :
 QLast Update : Tue Aug 17 02:29:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1	N-OCTANE	146.315	136.104 E3	7.0	93	0.00
2	N-DECANE	148.508	138.858 E3	6.5	93	0.00
3	N-DODECANE	150.730	140.706 E3	6.7	93	0.00
4	N-TETRADECANE	151.299	141.273 E3	6.6	93	0.00
5	N-HEXADECANE	149.261	139.432 E3	6.6	93	0.00
6	N-OCTADECANE	149.605	138.257 E3	7.6	93	0.00
7	N-EICOSANE	149.381	136.830 E3	8.4	92	0.00
8	N-DOCOSANE	145.120	132.215 E3	8.9	92	0.00
9 S	TETRACOSANE-d50 (SURROGATE)	120.062	109.085 E3	9.1	91	0.00
10	N-TETRACOSANE	139.018	126.198 E3	9.2	91	0.00
11	N-HEXACOSANE	132.065	119.820 E3	9.3	91	0.00
12	N-OCTACOSANE	125.139	113.193 E3	9.5	91	0.00
13	N-TRIACONTANE	118.437	105.585 E3	10.9	91	0.00
14	N-DOTRIACONTANE	110.652	97.539 E3	11.9	91	0.00
15	N-TETRATRIACONTANE	98.012	87.362 E3	10.9	92	0.00
16	N-HEXATRIACONTANE	91.581	82.602 E3	9.8	92	0.00
17	N-OCTATRIACONTANE	88.309	80.604 E3	8.7	92	0.00
18	N-TETRACONTANE	93.232	80.985 E3	13.1	92	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0