

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG081621\
 Data File : FG008676.D
 Signal(s) : FID1A.ch
 Acq On : 17 Aug 2021 01:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: Sample.e
 Quant Time: Aug 17 05:21:31 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG081621.M
 Quant Title :
 QLast Update : Tue Aug 17 02:29:16 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.754	5360520	44.648 ug/ml
Target Compounds			
2) N-DECANE	4.456	6821893	45.936 ug/ml
3) N-DODECANE	6.484	6923509	45.933 ug/ml
4) N-TETRADECANE	8.285	6925162	45.771 ug/ml
5) N-HEXADECANE	9.891	6807805	45.610 ug/ml
6) N-OCTADECANE	11.337	6738067	45.039 ug/ml
7) N-EICOSANE	12.651	6682978	44.738 ug/ml
8) N-DOCOSANE	13.853	6481735	44.665 ug/ml
10) N-TETRACOSANE	14.959	6217801	44.727 ug/ml
11) N-HEXACOSANE	15.984	5928498	44.891 ug/ml
12) N-OCTACOSANE	16.934	5641013	45.078 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG081621\
Data File : FG008676.D
Signal(s) : FID1A.ch
Acq On : 17 Aug 2021 01:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: Sample.e
Quant Time: Aug 17 05:21:31 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG081621.M
Quant Title :
QLast Update : Tue Aug 17 02:29:16 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

