

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082018\
 Data File : FG002132.D
 Signal(s) : FID1A.ch
 Acq On : 20 Aug 2018 12:04
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 20 15:49:10 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG073018.M
 Quant Title :
 QLast Update : Mon Jul 30 13:56:56 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.458	6339330	50.292 ug/ml
Target Compounds			
2) N-DECANE	4.209	7673410	52.066 ug/ml
3) N-DODECANE	6.222	7819146	52.048 ug/ml
4) N-TETRADECANE	8.013	7863407	51.696 ug/ml
5) N-HEXADECANE	9.611	7818842	51.333 ug/ml
6) N-OCTADECANE	11.052	7754453	50.554 ug/ml
7) N-EICOSANE	12.361	7579465	50.227 ug/ml
8) N-DOCOSANE	13.559	7371914	50.191 ug/ml
10) N-TETRACOSANE	14.662	7131440	49.740 ug/ml
11) N-HEXACOSANE	15.683	6878736	48.830 ug/ml
12) N-OCTACOSANE	16.633	6602188	47.323 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082018\
Data File : FG002132.D
Signal(s) : FID1A.ch
Acq On : 20 Aug 2018 12:04
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 20 15:49:10 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG073018.M
Quant Title :
QLast Update : Mon Jul 30 13:56:56 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

