

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082019\
 Data File : FG005660.D
 Signal(s) : FID1A.ch
 Acq On : 20 Aug 2019 16:41
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 20 22:46:32 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
 Quant Title :
 QLast Update : Thu Aug 08 16:03:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.748	2120807	16.600 ug/ml
Target Compounds			
12) N-OCTACOSANE	16.926	26913	0.178 ug/ml
13) N-TRIACONTANE	17.814	57622	0.381 ug/ml
14) N-DOTRIACONTANE	18.650	88873	0.664 ug/ml
15) N-TETRATRIACONTANE	19.466	97756	0.791 ug/ml
16) N-HEXATRIACONTANE	20.182	1264189	10.965 ug/ml
17) N-OCTATRIACONTANE	20.984	560352	5.367 ug/ml
18) N-TETRACONTANE	21.905	138663	1.333 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082019\
Data File : FG005660.D
Signal(s) : FID1A.ch
Acq On : 20 Aug 2019 16:41
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 20 22:46:32 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 16:03:53 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 14.748 min
Delta R.T.: -0.012 min
Response: 2120807
Conc: 16.60 ug/ml




#12 N-OCTACOSANE

R.T.: 16.926 min
Delta R.T.: -0.014 min
Response: 26913
Conc: 0.18 ug/ml




#13 N-TRIACONTANE

R.T.: 17.814 min
Delta R.T.: -0.016 min
Response: 57622
Conc: 0.38 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.650 min
Delta R.T.: -0.014 min
Response: 88873
Conc: 0.66 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.466 min
Delta R.T.: 0.017 min
Response: 97756
Conc: 0.79 ug/ml




#16 N-HEXATRIACONTANE

R.T.: 20.182 min
Delta R.T.: -0.012 min
Response: 1264189
Conc: 10.96 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.984 min
Delta R.T.: 0.033 min
Response: 560352
Conc: 5.37 ug/ml



#18 N-TETRACONTANE

R.T.: 21.905 min
Delta R.T.: 0.003 min
Response: 138663
Conc: 1.33 ug/ml