

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082219\
 Data File : FG005699.D
 Signal(s) : FID1A.ch
 Acq On : 22 Aug 2019 18:40
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 23 00:58:42 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
 Quant Title :
 QLast Update : Thu Aug 08 16:03:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.750	5999775	46.961 ug/ml
Target Compounds			
2) N-DECANE	4.449	7719114	54.239 ug/ml
3) N-DODECANE	6.480	7882383	53.806 ug/ml
4) N-TETRADECANE	8.281	7766917	52.396 ug/ml
5) N-HEXADECANE	9.888	7682345	51.147 ug/ml
6) N-OCTADECANE	11.334	7417339	49.718 ug/ml
7) N-EICOSANE	12.647	7270686	48.475 ug/ml
8) N-DOCOSANE	13.849	6995357	47.401 ug/ml
10) N-TETRACOSANE	14.955	6798961	46.781 ug/ml
11) N-HEXACOSANE	15.978	6622571	46.140 ug/ml
12) N-OCTACOSANE	16.929	6515884	43.103 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082219\
Data File : FG005699.D
Signal(s) : FID1A.ch
Acq On : 22 Aug 2019 18:40
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 23 00:58:42 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 16:03:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

