

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082419\
Data File : FG005716.D
Signal(s) : FID1A.ch
Acq On : 24 Aug 2019 10:22
Operator : DD\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 26 01:57:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 16:03:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
13) N-TRIACONTANE	17.814	45492	0.301 ug/ml
14) N-DOTRIACONTANE	18.651	48903	0.365 ug/ml
15) N-TETRATRIACONTANE	19.435	38283	0.310 ug/ml
16) N-HEXATRIACONTANE	20.205	169933	1.474 ug/ml
17) N-OCTATRIACONTANE	20.934	190567	1.825 ug/ml
18) N-TETRACONTANE	21.892	108157	1.040 ug/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082419\
Data File : FG005716.D
Signal(s) : FID1A.ch
Acq On : 24 Aug 2019 10:22
Operator : DD\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 26 01:57:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 16:03:53 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um






#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 0.000 min
Exp R.T.: 14.760 min
Response: 0
Conc: N.D.




#13 N-TRIACONTANE

R.T.: 17.814 min
Delta R.T.: -0.015 min
Response: 45492
Conc: 0.30 ug/ml




#14 N-DOTRIACONTANE

R.T.: 18.651 min
Delta R.T.: -0.013 min
Response: 48903
Conc: 0.37 ug/ml




#15 N-TETRATRIACONTANE

R.T.: 19.435 min
Delta R.T.: -0.015 min
Response: 38283
Conc: 0.31 ug/ml




#16 N-HEXATRIACONTANE

R.T.: 20.205 min
Delta R.T.: 0.011 min
Response: 169933
Conc: 1.47 ug/ml



#17 N-OCTATRIACONTANE

R.T.: 20.934 min
Delta R.T.: -0.017 min
Response: 190567
Conc: 1.83 ug/ml



#18 N-TETRACONTANE

R.T.: 21.892 min
Delta R.T.: -0.010 min
Response: 108157
Conc: 1.04 ug/ml