

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082419\
 Data File : FG005718.D
 Signal(s) : FID1A.ch
 Acq On : 24 Aug 2019 11:22
 Operator : DD\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 26 01:58:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
 Quant Title :
 QLast Update : Thu Aug 08 16:03:53 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.745	6168855	48.285 ug/ml
Target Compounds			
2) N-DECANE	4.447	7729966	54.315 ug/ml
3) N-DODECANE	6.476	7928822	54.123 ug/ml
4) N-TETRADECANE	8.278	7871946	53.104 ug/ml
5) N-HEXADECANE	9.884	7839106	52.191 ug/ml
6) N-OCTADECANE	11.330	7607777	50.995 ug/ml
7) N-EICOSANE	12.642	7463467	49.760 ug/ml
8) N-DOCOSANE	13.844	7187867	48.705 ug/ml
10) N-TETRACOSANE	14.951	6992745	48.115 ug/ml
11) N-HEXACOSANE	15.973	6798477	47.366 ug/ml
12) N-OCTACOSANE	16.923	6682385	44.204 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082419\
Data File : FG005718.D
Signal(s) : FID1A.ch
Acq On : 24 Aug 2019 11:22
Operator : DD\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 26 01:58:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080819.M
Quant Title :
QLast Update : Thu Aug 08 16:03:53 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

