

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082622\
 Data File : FG010435.D
 Signal(s) : FID1A.ch
 Acq On : 26 Aug 2022 21:36
 Operator : YP\AJ
 Sample : 5 TRPH STD
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 5 TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/29/2022
 Supervised By :Ankita Jodhani 08/29/2022

Integration File: autoint1.e
 Quant Time: Aug 26 23:47:19 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
 Quant Title :
 QLast Update : Fri Aug 26 23:38:53 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.785	643323	5.257 ug/ml
Target Compounds			
1) N-OCTANE	1.816	661902	5.260 ug/ml
2) N-DECANE	4.316	666727	5.202 ug/ml
3) N-DODECANE	6.495	682063	5.182 ug/ml
4) N-TETRADECANE	8.325	705173	5.210 ug/ml
5) N-HEXADECANE	9.932	730748	5.215 ug/ml
6) N-OCTADECANE	11.375	746012	5.225 ug/ml
7) N-EICOSANE	12.686	749030	5.246 ug/ml
8) N-DOCOSANE	13.884	731315	5.248 ug/ml
10) N-TETRACOSANE	14.988	726003	5.240 ug/ml
11) N-HEXACOSANE	16.010	709241	5.228 ug/ml
12) N-OCTACOSANE	16.960	694749	5.254 ug/ml
13) N-TRIACONTANE	17.849	694111	5.357 ug/ml
14) N-DOTRIACONTANE	18.682	719791	5.630 ug/ml
15) N-TETRATRIACONTANE	19.463	688293	5.564 ug/ml
16) N-HEXATRIACONTANE	20.204	623263	5.370 ug/ml
17) N-OCTATRIACONTANE	20.915	595298	5.183 ug/ml
18) N-TETRACONTANE	21.749	581204	4.935 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082622\
Data File : FG010435.D
Signal(s) : FID1A.ch
Acq On : 26 Aug 2022 21:36
Operator : YP\AJ
Sample : 5 TRPH STD
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
5 TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/29/2022
Supervised By :Ankita Jodhani 08/29/2022

Integration File: autoint1.e
Quant Time: Aug 26 23:47:19 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG082622.M
Quant Title :
QLast Update : Fri Aug 26 23:38:53 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

