

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG08262

Data File : FG016389.D

Signal(s) : FID1A.ch

Acq On : 26 Aug 2025 09:24

Sample : 50 PPM TRPH STD

Misc :

ALS Vial : 3 Sample Multiplier: 1

Reviewed By :Yogesh Patel 08/27/2025

Supervised By :mohammad ahmed 08/28/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080525.M

Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.114	2.075	2.224	PB	360043	4629356	84.52%	5.240%
2	4.663	4.605	4.772	BB	486689	5082759	92.79%	5.753%
3	6.847	6.789	6.937	BB	523722	5187357	94.70%	5.871%
4	8.688	8.600	8.779	BB	489341	5250837	95.86%	5.943%
5	10.308	10.255	10.390	BB	485521	5322326	97.17%	6.024%
6	11.764	11.657	11.832	BB	479317	5448656	99.47%	6.167%
7	13.085	12.987	13.140	BV	464887	5477414	100.00%	6.200%
8	14.293	14.258	14.362	VB	431773	5330420	97.32%	6.033%
9	15.199	15.154	15.228	BV	337395	4687779	85.58%	5.306%
10	15.405	15.370	15.484	VB	425527	5326232	97.24%	6.028%
11	16.434	16.245	16.485	PV	420231	5267145	96.16%	5.962%
12	17.392	17.297	17.452	BV	387509	5249376	95.84%	5.942%
13	18.285	18.202	18.325	VV	395847	5452836	99.55%	6.172%
14	19.125	19.039	19.174	VV	383679	5458221	99.65%	6.178%
15	19.914	19.857	19.951	VV	337699	4833498	88.24%	5.471%
16	20.658	20.625	20.698	BV	271563	3930836	71.76%	4.449%
17	21.449	21.410	21.527	BV	173452	3314612	60.51%	3.752%
18	22.457	22.374	22.572	VV	117974	3101208	56.62%	3.510%
Sum of corrected areas:						88350869		

FG080525.M Wed Aug 27 05:31:11 2025