

rteres

Instrument :

FID_G

LabSampleId :

50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG08262

Data File : FG016393.D

Signal(s) : FID1A.ch

Acq On : 26 Aug 2025 11:37

Sample : 50 PPM TRPH STD

Misc :

ALS Vial : 55 Sample Multiplier: 1

Reviewed By :Yogesh Patel 08/27/2025

Supervised By :mohammad ahmed 08/28/2025

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080525.M

Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.114	2.076	2.237	VB	402889	5138697	91.48%	5.780%
2	4.662	4.629	4.762	BB	510745	5325195	94.80%	5.989%
3	6.846	6.791	6.941	BB	531633	5395272	96.05%	6.068%
4	8.688	8.621	8.769	BB	523172	5451257	97.05%	6.131%
5	10.308	10.269	10.386	BB	492621	5415864	96.41%	6.091%
6	11.763	11.647	11.839	BB	482791	5508889	98.07%	6.196%
7	13.084	12.996	13.134	BV	486304	5617243	100.00%	6.318%
8	14.292	14.219	14.374	BB	455998	5449593	97.02%	6.129%
9	15.197	15.157	15.264	BV	355044	4733827	84.27%	5.324%
10	15.404	15.364	15.471	VB	424266	5417408	96.44%	6.093%
11	16.432	16.344	16.504	BB	405866	5256693	93.58%	5.912%
12	17.390	17.177	17.452	BV	391831	5276124	93.93%	5.934%
13	18.283	18.203	18.329	PV	380301	5280978	94.01%	5.940%
14	19.122	18.986	19.174	VV	367465	5145498	91.60%	5.787%
15	19.911	19.621	19.951	BV	327258	4612057	82.11%	5.187%
16	20.656	20.626	20.696	BV	265046	3774392	67.19%	4.245%
17	21.447	21.411	21.511	BV	176120	3164597	56.34%	3.559%
18	22.454	22.357	22.569	BB	112393	2945389	52.43%	3.313%
Sum of corrected areas:						88908972		

FG080525.M Wed Aug 27 05:37:10 2025