

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082823\
 Data File : FG012799.D
 Signal(s) : FID1A.ch
 Acq On : 28 Aug 2023 16:25
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Aug 29 04:14:27 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080223.M
 Quant Title :
 QLast Update : Wed Aug 02 18:57:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.082	5297111	47.352 ug/ml
Target Compounds			
2) N-DECANE	4.600	6464300	50.885 ug/ml
3) N-DODECANE	6.774	6508876	50.674 ug/ml
4) N-TETRADECANE	8.606	6482036	50.318 ug/ml
5) N-HEXADECANE	10.219	6471154	49.882 ug/ml
6) N-OCTADECANE	11.667	6472069	49.345 ug/ml
7) N-EICOSANE	12.980	6275682	48.793 ug/ml
8) N-DOCOSANE	14.182	6058171	47.900 ug/ml
10) N-TETRACOSANE	15.288	5965514	47.171 ug/ml
11) N-HEXACOSANE	16.311	5773189	46.120 ug/ml
12) N-OCTACOSANE	17.262	5626265	45.778 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG082823\
Data File : FG012799.D
Signal(s) : FID1A.ch
Acq On : 28 Aug 2023 16:25
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Aug 29 04:14:27 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080223.M
Quant Title :
QLast Update : Wed Aug 02 18:57:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

