

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG083018\
 Data File : FG002208.D
 Signal(s) : FID1A.ch
 Acq On : 30 Aug 2018 14:56
 Operator : AJ\SJ
 Sample : FG083018ICV
 Misc :
 ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Aug 30 15:28:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:49:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.429	5978240	47.385 ug/ml
Target Compounds			
1) N-OCTANE	2.117	5705350	48.224 ug/ml
2) N-DECANE	4.189	5816571	48.457 ug/ml
3) N-DODECANE	6.200	6033617	47.944 ug/ml
4) N-TETRADECANE	7.991	6266640	47.735 ug/ml
5) N-HEXADECANE	9.587	6480628	47.612 ug/ml
6) N-OCTADECANE	11.025	6653956	47.371 ug/ml
7) N-EICOSANE	12.333	6722048	47.316 ug/ml
8) N-DOCOSANE	13.530	6759050	47.359 ug/ml
10) N-TETRACOSANE	14.632	6758352	47.274 ug/ml
11) N-HEXACOSANE	15.652	6768043	47.139 ug/ml
12) N-OCTACOSANE	16.601	6781091	46.347 ug/ml
13) N-TRIACONTANE	17.489	6804018	44.855 ug/ml
14) N-DOTRIACONTANE	18.321	6606487	44.793 ug/ml
15) N-TETRATRIACONTANE	19.107	6394925	46.352 ug/ml
16) N-HEXATRIACONTANE	19.849	6022789	46.963 ug/ml
17) N-OCTATRIACONTANE	20.558	5326062	47.376 ug/ml
18) N-TETRACONTANE	21.377	5156619	47.517 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG083018\
Data File : FG002208.D
Signal(s) : FID1A.ch
Acq On : 30 Aug 2018 14:56
Operator : AJ\SJ
Sample : FG083018ICV
Misc :
ALS Vial : 58 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Aug 30 15:28:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG083018.M
Quant Title :
QLast Update : Thu Aug 30 14:49:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

