

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG083123\
 Data File : FG012846.D
 Signal(s) : FID1A.ch
 Acq On : 31 Aug 2023 11:52
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/01/2023
 Supervised By :mohammad ahmed 09/01/2023

Integration File: autoint1.e
 Quant Time: Sep 01 00:15:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080223.M
 Quant Title :
 QLast Update : Wed Aug 02 18:57:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.083	5551735	49.628 ug/mlm
Target Compounds			
1) N-OCTANE	2.078	6572484	53.383 ug/ml
2) N-DECANE	4.600	6896703	54.289 ug/ml
3) N-DODECANE	6.774	6876213	53.534 ug/ml
4) N-TETRADECANE	8.607	6800963	52.794 ug/ml
5) N-HEXADECANE	10.221	6764842	52.145 ug/ml
6) N-OCTADECANE	11.668	6757908	51.525 ug/ml
7) N-EICOSANE	12.981	6567240	51.060 ug/ml
8) N-DOCOSANE	14.183	6355305	50.250 ug/ml
10) N-TETRACOSANE	15.290	6263046	49.523 ug/ml
11) N-HEXACOSANE	16.313	6080929	48.579 ug/ml
12) N-OCTACOSANE	17.264	5959993	48.493 ug/ml
13) N-TRIACONTANE	18.152	5789218	48.882 ug/ml
14) N-DOTRIACONTANE	18.987	5287136	47.079 ug/ml
15) N-TETRATRIACONTANE	19.769	4582497	48.378 ug/mlm
16) N-HEXATRIACONTANE	20.510	3848002	47.419 ug/mlm
17) N-OCTATRIACONTANE	21.266	3403467	47.130 ug/ml
18) N-TETRACONTANE	22.215	3478627	49.924 ug/mlm

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG083123\
Data File : FG012846.D
Signal(s) : FID1A.ch
Acq On : 31 Aug 2023 11:52
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/01/2023
Supervised By :mohammad ahmed 09/01/2023

Integration File: autoint1.e
Quant Time: Sep 01 00:15:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080223.M
Quant Title :
QLast Update : Wed Aug 02 18:57:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

