

rteres

Instrument :
FID_G
LabSampleId :
50 PPM TRPH STD

Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 09/01/2023
Supervised By :mohammad ahmed 09/01/2023

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG012841.D
Data File : FG012841.D
Signal(s) : FID1A.ch
Acq On : 31 Aug 2023 09:24
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG080223.M
Title :

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	2.081	2.038	2.151	BB	703424	7150584	94.47%	6.286%
2	4.601	4.560	4.645	BB	746639	7550387	99.75%	6.638%
3	6.775	6.729	6.818	BB	741972	7569477	100.00%	6.654%
4	8.607	8.568	8.648	BB	729431	7516140	99.30%	6.607%
5	10.221	10.180	10.268	BB	685695	7487459	98.92%	6.582%
6	11.668	11.623	11.703	BB	658035	7477999	98.79%	6.574%
7	12.982	12.936	13.031	BB	612103	7257161	95.87%	6.380%
8	14.184	14.146	14.233	VB	583982	7007568	92.58%	6.160%
9	15.085	15.035	15.118	BV	463502	6122109	80.88%	5.382%
10	15.291	15.218	15.345	BB	533136	6892467	91.06%	6.059%
11	16.312	16.261	16.366	BB	514150	6677905	88.22%	5.871%
12	17.265	17.116	17.319	BB	487229	6532317	86.30%	5.743%
13	18.153	18.089	18.191	BB	470748	6302496	83.26%	5.541%
14	18.987	18.928	19.036	BB	402552	5741097	75.85%	5.047%
15	19.771	19.651	19.808	BB	346660	4949962	65.39%	4.352%
16	20.509	20.205	20.559	BB	278913	4172386	55.12%	3.668%
17	21.268	21.197	21.313	BB	208469	3714136	49.07%	3.265%
18	22.218	22.131	22.300	BBA	147161	3630800	47.97%	3.192%

Sum of corrected areas: 113752450

FG080223.M Fri Sep 01 01:20:00 2023