

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_G\Data\FG083123\  
Data File : FG012844.D  
Signal(s) : FID1A.ch  
Acq On : 31 Aug 2023 10:53  
Operator : YP\AJ  
Sample : 04221-01 5X  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
FID\_G  
ClientSampleId :  
AU-5-082923

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 09/01/2023  
Supervised By :mohammad ahmed 09/01/2023

Integration File: autoint1.e  
Quant Time: Sep 01 00:15:29 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_G\Method\FG080223.M  
Quant Title :  
QLast Update : Wed Aug 02 18:57:31 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um

| Compound                      | R.T.   | Response | Conc Units   |
|-------------------------------|--------|----------|--------------|
| -----                         |        |          |              |
| System Monitoring Compounds   |        |          |              |
| 9) S TETRACOSANE-d50 (SURR... | 15.078 | 426155   | 3.810 ug/mlm |

Target Compounds  
-----

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_G\Data\FG083123\  
Data File : FG012844.D  
Signal(s) : FID1A.ch  
Acq On : 31 Aug 2023 10:53  
Operator : YP\AJ  
Sample : 04221-01 5X  
Misc :  
ALS Vial : 7 Sample Multiplier: 1

**Instrument :**

FID\_G

**ClientSampleId :**

AU-5-082923


**Manual Integrations****APPROVED**


Reviewed By :Yogesh Patel 09/01/2023

Supervised By :mohammad ahmed 09/01/2023

Integration File: autoint1.e  
Quant Time: Sep 01 00:15:29 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_G\Method\FG080223.M  
Quant Title :  
QLast Update : Wed Aug 02 18:57:31 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 1uL  
Signal Phase : Rxi-1ms  
Signal Info : 20mx0.18mmx0.18um





#9 TETRACOSANE-d50 (SURROGATE)

R.T.: 15.078 min  
Delta R.T.: -0.004 min  
Response: 426155  
Conc: 3.81 ug/ml

Instrument :

FID\_G

ClientSampleId :

AU-5-082923

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 09/01/2023

Supervised By :mohammad ahmed 09/01/2023