

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090518\
 Data File : FG002248.D
 Signal(s) : FID1A.ch
 Acq On : 05 Sep 2018 17:07
 Operator : AJ\SJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Sep 06 01:44:23 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG083018.M
 Quant Title :
 QLast Update : Thu Aug 30 14:49:31 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	14.426	5891672	46.699 ug/ml
Target Compounds			
2) N-DECANE	4.188	6624070	55.185 ug/ml
3) N-DODECANE	6.199	6801964	54.049 ug/ml
4) N-TETRADECANE	7.989	6949197	52.934 ug/ml
5) N-HEXADECANE	9.587	7044519	51.755 ug/ml
6) N-OCTADECANE	11.025	7080207	50.405 ug/ml
7) N-EICOSANE	12.332	6958172	48.978 ug/ml
8) N-DOCOSANE	13.528	6796384	47.621 ug/ml
10) N-TETRACOSANE	14.629	6600046	46.167 ug/ml
11) N-HEXACOSANE	15.650	6406713	44.623 ug/ml
12) N-OCTACOSANE	16.598	6202793	42.394 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090518\
Data File : FG002248.D
Signal(s) : FID1A.ch
Acq On : 05 Sep 2018 17:07
Operator : AJ\SJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Sep 06 01:44:23 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG083018.M
Quant Title :
QLast Update : Thu Aug 30 14:49:31 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

