

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090623\
 Data File : FG012857.D
 Signal(s) : FID1A.ch
 Acq On : 06 Sep 2023 13:19
 Operator : YP\AJ
 Sample : 50 PPM TRPH STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 FID_G
 ClientSampleId :
 50 PPM TRPH STD

Integration File: autoint1.e
 Quant Time: Sep 07 01:42:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090123.M
 Quant Title :
 QLast Update : Thu Aug 31 16:39:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1uL
 Signal Phase : Rxi-1ms
 Signal Info : 20mx0.18mmx0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S TETRACOSANE-d50 (SURR...	15.087	6386408	57.136 ug/ml
Target Compounds			
2) N-DECANE	4.599	7909809	58.389 ug/ml
3) N-DODECANE	6.775	7913991	58.472 ug/ml
4) N-TETRADECANE	8.609	7835122	58.412 ug/ml
5) N-HEXADECANE	10.222	7810149	58.268 ug/ml
6) N-OCTADECANE	11.670	7778433	58.105 ug/ml
7) N-EICOSANE	12.984	7532630	57.961 ug/ml
8) N-DOCOSANE	14.186	7284376	57.933 ug/ml
10) N-TETRACOSANE	15.292	7173432	57.848 ug/ml
11) N-HEXACOSANE	16.315	6983884	58.126 ug/ml
12) N-OCTACOSANE	17.266	6816067	58.181 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_G\Data\FG090623\
Data File : FG012857.D
Signal(s) : FID1A.ch
Acq On : 06 Sep 2023 13:19
Operator : YP\AJ
Sample : 50 PPM TRPH STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
FID_G
ClientSampleId :
50 PPM TRPH STD

Integration File: autoint1.e
Quant Time: Sep 07 01:42:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_G\Method\FG090123.M
Quant Title :
QLast Update : Thu Aug 31 16:39:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1uL
Signal Phase : Rxi-1ms
Signal Info : 20mx0.18mmx0.18um

